

IMCW2 Assembly #400

Configuration Worksheet

Customer Name _____
 Customer Number _____
 Customer PO# _____
 Customer PO Line _____
 Customer P/N _____
 Dresser P/N _____

1. Select a Meter Series this unit is for:

- Series A 1.5M – 16M Series B
 Series A 23M – 102M Romet TC RM2/3/5M Romet TC RM7/11/16/23M

2. Select Corrector type:

- Pressure (P-only) Temperature (T-only) Pressure & Temperature (P&T)

3. Select Pressure transducer type and the Pressure range if unit is P-only or P & T:

- Gauge: 30 psi (2 bar) 180 psi (12 bar) 1480 psi (100 bar)
 Absolute: 30 psi (2 bar) 180 psi (12 bar) 1480 psi (100 bar)

4. Select Pressure Connection type:

- Series B: Internal (for 8C – 16M) External (for 8C-56M; 23M232; 102M; G16-G1000; HPC Meters)
 Series A & Romet: External only

5. Select Temperature Probe if unit is T-only or P&T:

- Internal – Series A (1.5-16M) External – Series A
 Internal – Series B (8C-16M) External – Series B
 Internal – Romet (2M-23M TC)

For external temperature, select a probe & length of cable:

- T1 (8-1/4" 5-ft) T2 (8-1/4" 10-ft) T3 (2" 5-ft) T4 (2" 10-ft)

Note: All cables are armored.

Note: Choose T1 or T2 if the unit is for 102M, 23-56M175, 1M/3M740 & 1480, 5M/7M/11M1480 meters.

6. Select Output Connection type and Cable Length if required:

- MS Circular (single) MS Circular (dual)
 None (plug)
 Cable Gland (single) No Cable 5' 30' 50'
 Cable Gland (dual) No Cable 5' 30' 50'
 Conduit No Cable 5' 30' 50' None (plug)

7. Select Battery Pack type:

- Alkaline Lithium

8. Select Software Configuration type:

- Default Imperial Default Metric
 Custom (Consult factory to set up custom configuration)

Dresser Measurement
16240 Port Northwest Drive
Houston, TX 77041
T: 1-800-521-1114
F: 1-800-335-5224

The Dresser Logo and all Trademarks containing the term "Dresser" are the property of Dresser, LLC, a subsidiary of Baker Hughes.

